

CERTIFICATE OF ANALYSIS
FP: Fibrinogen Reagent 5 x 2 mL



Temp.: TF0326.07

Attachment 11 DQ6E0A.03

6E36CA0.03

REF 5138085-RUO
LOT 6E36CA0.03
⌚ 2025-12-31

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	grey		confirmed	OK
	Vial type	clear glass vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material		confirmed	OK
	Appearance after reconstitution	clear liquid		confirmed	OK
Recovery of controls Fibrinogen	Control N LOT 1P31C01	2.67	-	4.01 g/L	3.68 g/L OK
	Control A LOT 3P42C00	0.93	-	1.39 g/L	1.25 g/L OK

Prepared by

Smilja DURANOVIĆ
Smilja Duranovic

15. JAN. 2025

Date

Approved by

Stefan BRUCKBERGER
Stefan Bruckberger

16. JAN. 2025

Date