

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2K00.02
---	--	---------------------------------

F VIII Inhibitorplasma 5x1 mL

REF	5159008-RUO
LOT	2K42C00.03
	2026-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability	FVIII Inhibitor clotting, Siron LS	16.3 – 27.1 BE/mL	26.0	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

Durchführung von
Prepared by

Lukas GRUBELNIG 

09. SEP. 2025
Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

09. SEP. 2025
Datum / Date: