



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ4E00.04

Att.03

4E41C00.02

## F XI Mangelplasma nativ, 5x1mL

REF	5184004-RUO
LOT	4E41C00.02
⌚	2026-03-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Cap color	blue	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
Appearance after reconstitution	buff colored liquid without visible particles	confirmed	OK	
Virologie / Virology	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab  The testing methods applied were FDA-approved or CE marked.	Negative Negative Positive	Negative Negative Positive	OK
Kontrollwiederfindung F XI am Ceveron® Alpha	<b>Control N</b> LOT 1P31C00	74.3 - 123.8 %	90.1 %	OK
Recovery of controls F XI on Ceveron® Alpha	<b>Control A</b> LOT 3P42C00	27.0 - 45.0 %	35.3 %	OK

Durchgeführt von

Prepared by

Maylis RAUSCHER

19. JUNI 2024

Datum / Date:

Freigabe von

Approved by

Lavdrim NASUFI

19. JUNI 2024

Datum / Date: