## Coagulation Reference 5x1 ml



Attachment 04

DQ1R00.02

DQINOON

1R45C00.02

REF

5220110-RUO

LOT

1R45C00.02

2

2026-08-31

Analytical Procedures		Requirement			Result			
	Cap color		green		confir	med	ок	
visual inspection	Vial type	clear glass vial		confir	med	ок		
	Vial label	according to packaging SOP		confir	med	ок		
	Vial barcode	according to packaging SOP			confir	med	ОК	
	Box label	according to packaging SOP			confirmed		ок	
	Box barcode	according to packaging SOP		confirmed		ок		
	Package insert	present		confirmed		ок		
	Batch table	present		confir	med	ок		
	Appearance before reconstitution	buff colored plug of lyophilized material		confir	med	ок		
	Appearance after reconstitution	clear liquid		confirmed		ок		
	HBsAg / HBsAg	Negative		Negative				
Virology	HIV-Ak / HIV-Ab	Negative		Negative				
Self (Art	HCV-AK / HCV-Ab	Negative		Negative		OK		
The testing methods applied were FDA-approved or CE marked.								
	Control N	2.7		4.0		2.5		
Recovery of controls	LOT 1P31C01	2.7	-	4.0	g/l	3.5	g/l	ОК
Fibrinogen	Control A	0.0			- //	4.0	- //	01/
	LOT 3P42C00	0.9	-	1.4	g/l	1.2	g/l	ОК

Prepared by	Maylis HAUSCHER	2 7. AUG. 2024			
		Date			
Approved by	FIKTET NASUFI Filmet Wyul	2 7. AUG. 2024			
	,	Date			