Coagulation Reference 5x1 ml



Attachment 04

DQ1R00.02

1R04C02.02

REF

5220110-RUO

1R04C02.02

8

2025-08-31

Analytical Procedures		Requirement			Result			
	Cap color		green		confirm	med	ок	
visual inspection	Vial type	С	clear glass vial		confir	ned	ок	
	Vial label	according to packaging SOP			confir	med	ОК	
	Vial barcode	according to packaging SOP			confir	med	ок	
	Box label	according to packaging SOP			confir	med	ОК	
	Box barcode	according to packaging SOP		confir	med	ОК		
	Package insert	present		confir	med	ОК		
	Batch table	present		confir	med	ОК		
	Appearance before reconstitution	buff colored plug of lyophilized material		confir	med	ок		
	Appearance after reconstitution	clear liquid			confir	med	ОК	
	HBsAg / HBsAg	Negative		Negative		ок		
Virology	HIV-Ak / HIV-Ab	Negative			Negative			
ANALYSIS .	HCV-AK / HCV-Ab	Negative			Negative			
The testing methods applied were FDA-approved or CE marked.								
Recovery of controls	Control N	71.2	-	118.8	%	95.3	%	ок
	LOT 1P31C00	71.3					70	
FVII clotting	Control A	33.0	12	55.0	%	42.6 %	%	ок
	LOT 3P41C00			JJ.U			70	UK

Pre	pared	by
	puice	,

LUGARHONSKICZ

2 8. FEB. 2024

Approved by

Maylis RAUSCHER

2 9. FEB. 2024

Date