



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0W00.01

Technochrom AT III Kit ~ 100 T.

REF 5340224-RUO

LOT 0W33C00.07

🕒 2025-09-30

Inhalt Content	Reagenz Reagent	LOT	🕒
2x25 mL	NaCl Solution 0,9 %	9524C00	2027-10-31
1x	ATIII Substrate Th-1 10 µmol	8B21C00	2025-09-30
1x	Technochrom ATIII Reagent A2 20 µmol	7S21C00	2025-10-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Ceveron® s100 AT-factor IIa based	CCN LOT 1P11C01 78.0 – 130.0 %	105.8 OK
	CCA LOT 3P23C00 36.8 – 61.3 %	47.7 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

Durchführung von
Prepared by

Maylis RAUSCHER

27. SEP. 2024

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

27. SEP. 2024

Datum / Date: