

CERTIFICATE OF ANALYSIS

Technochrom anti Xa Kit



Attachment 06 DQ0800.01

Temp.: TF0327.07

0841C00.02

REF	5340250-RUO
LOT	0841C00.02
	2026-03-31

Content	volume	Reagent	LOT	
1x	4 mL	Technochrom anti-Xa Fxa Reagent	7K31C01	2026-03-31
1x	4 mL	Technochrom anti-Xa SUB	8A31C01	2026-03-31
1x	20 mL	Technochrom anti-Xa BUF	9R31C00	2028-01-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Edoxaban Cont L LOT P741C00	23 - 43 ng/mL	33 ng/mL	OK
	Edoxaban Cont M LOT P841C00	95 - 158 ng/mL	127 ng/mL	OK

Prepared by Lucas HOLOWICZ

18. MRZ. 2024
Date

Approved by Michael KAFKA

18. MRZ. 2024
Date