## **CERTIFICATE OF ANALYSIS Technochrom anti Xa Kit**



Attachment 06

DQ0800.01

0842C00.02

REF

5340250-RUO

LOT

0842C00.02

23

2026-03-31

Content	volume	Reagent	LOT	8	
1x	4 mL	Technochrom anti-Xa Fxa Reagent	7K41C00	2026-04-30	
1x	4 mL	Technochrom anti-Xa SUB	8A41C00	2026-03-31	
1x	20 mL	Technochrom anti-Xa BUF	9R41C00	2029-03-31	

<b>Analytical Procedures</b>		Requirement			Result			
	Content	according to packaging SOP		g SOP	confi	rmed	ОК	
	Vial label	according to packaging SOP		g SOP	confirmed		ОК	
a destal de accepto o	Vial barcode	according to packaging SOP			g SOP	confirmed		ОК
visual inspection	Box label	according to packaging SOP			g SOP	confirmed		ОК
	Box barcode	according to packaging SOP			confirmed		ОК	
	Package insert	present			confirmed		ок	
Daggiewi of saskuala	Edoxaban Cont L LOT P722C01	22.0	-	40.0	ng/mL	36.0	ng/mL	ОК
Recovery of controls	Edoxaban Cont M LOT P841C00	95.0	-	158.0	ng/mL	128.0	ng/mL	ОК

Pre	pared	bv
110	Duica	Uy

Maylis RAUSCHER

2 A. MAI 2024

Date

Approved by

Michael KAFKA

2 8. MAI 2024

Date