

technoclone TC	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0K00.03 Att.12 0K51C00.02
--------------------------------------	--	--

TECHNOCHROM® Protein C Reagent Kit

REF	5341013-RUO
LOT	0K51C00.02
	2026-05-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	1mL	Reference Standard PC 1	2N51C00	2027-05-31
1x	1mL	Reference Standard PC 2	2O51C00	2027-05-31
1x	1mL	Reference Standard PC 3	2P51C00	2027-05-31
1x	60mL	Protein C Puffer	9O41C00	2029-03-31
3x	1mL	Protac	7E41C00	2026-05-31
3x	1mL	Substrate for Protein C	8R43C00	2026-11-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement		Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
Kontrollwiederfindung Recovery of controls	Control N LOT 1P42C00	69.0	- 115.0 %	94.6 %	OK
	Control A LOT 3P42C01	12.8	- 21.3 %	19.1 %	OK

Durchgeföhrt von Prepared by	Lavdrin NASUFI	01. JULI 2025 Datum / Date:
Freigabe von Approved by	Maylis RAUSCHER	01. JULI 2025 Datum / Date: