

technoclone TC	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0K00.03 Att.12 0K52C00.02
--------------------------------------	--	--

TECHNOCHROM® Protein C Reagent Kit

REF

LOT

5341013-RUO

0K52C00.02

2026-11-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	1mL	Reference Standard PC 1	2N51C00	2027-05-31
1x	1mL	Reference Standard PC 2	2O51C00	2027-05-31
1x	1mL	Reference Standard PC 3	2P51C00	2027-05-31
1x	60mL	Protein C Puffer	9O51C00	2030-05-31
3x	1mL	Protac	7E52C00	2027-06-30
3x	1mL	Substrate for Protein C	8R43C00	2026-11-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Kontrollwiederfindung Recovery of controls	Control N LOT 1P42C00	69.0 - 115.0 %	89.4 %	OK
	Control A LOT 3P51C00	15.0 - 25.0 %	22.1 %	OK

Durchgeführt von
Prepared by

Lavdini NASUFI

21. AUG. 2025
Datum / Date:

Freigabe von
Approved by

Lukas GRUBELNIG

21. AUG. 2025
Datum / Date: