

CERTIFICATE OF ANALYSIS

Platelet Poor Plasma 5 x 2 ml



Attachment 10 DQ0L00.04

1L52C00.02

Temp.: TP0326.07

REF 5343022-RUO
 LOT 1L52C00.02
 2027-03-31

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	brown	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology		HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	OK
The testing methods applied were FDA-approved or CE marked.				
Recovery of control	LA Ratio	0.77 - 1.28	1.04	OK

Prepared by

Lukas GRUBELNIG *Lukas Grubelnig*

14. JULI 2025

Date

Approved by

FIKREI NASUFI *Fikrei Nasufi*

15. JULI 2025

Date