

CERTIFICATE OF ANALYSIS TECHNOCHROM C1-INH



Attachment 16 DQ0J00.03

0J55C00.04

Temp.: TP0327.07

REF 5345003-RUO
LOT 0J55C00.04
2027-02-28

Content	volume	Reagent	LOT	
1x	1mL	Coagulation Control N f. C1-INH	1P51CC1H	2027-03-31
1x	1mL	Coagulation Control A f. C1-INH	3P52CC1H	2027-03-31
1x	see label	Coagulation Reference f. C1-INH	1R53CC1H	2027-03-31
1x	3mL	Substrate C1-1	8C51C00	2027-02-28
1x	3mL	C1-Esterase	7B51C00	2027-07-31
1x	25mL	Sample Buffer A	9M51C00	2030-03-31
1x	20mL	Reaction Buffer B	9N51C00	2030-03-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N LOT 1P51CC1H	88.4 - 147.4 %	117.1 %	OK
	Control A LOT 3P52CC1H	39.7 - 66.1 %	50.2 %	OK

Prepared by

Lavdrin NASUFI

24. SEP. 2025

Date

Approved by

Lukas GRUBELNIG

24. SEP. 2025

Date