

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00.02
---	---	-----------

TECHNOCHROM FXIII 3 x 3 ml

REF 5360010-RUO


LOT 0Z53C00.01

 2026-06-30

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	2503100/1	2026-06-30
3 x 3 mL	FXIII Detection Reagent	2503100/2	2026-06-30
3 x 3 mL	FXIII NADPH	2503100/3	2026-06-30
3 x 1 mL	FXIII Inhibitor Reagent	2503100/4	2026-06-30
1 x 6 mL	Stabilizer solution	2503100/5	2026-06-30

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P42C00	78.8 – 131.3 % 118.0
	Coagulation Control A Lot: 3P42C01	30.0 – 50.0 % 39.9

Durchführung von
Prepared by


Lucas MOLOVICZ

06. AUG. 2025

Datum / Date:

Freigabe von
Approved by


Lukas GRUBELNIG

06. AUG. 2025

Datum / Date: