

# CERTIFICATE OF ANALYSIS

## Technochrom FXIII



Attachment 02 DQ0700.01

Temp.: TP0327.07

0Z56C00.02

REF 5360010-RUO  
 LOT 0Z56C00.02  
 2026-11-30

Content	volume	Reagent	LOT	2026-11-30
3x	3mL	FXIII Activator Reagent	2506100/1	2026-11-30
3x	3mL	FXIII Detection Reagent	2506100/2	2026-11-30
3x	3mL	FXIII NADPH Solution	2506100/3	2026-11-30
3x	1mL	FXIII Inhibitor Reagent	2506100/4	2026-11-30
1x	6mL	FXIII Stabilizer Solution	2506100/5	2026-11-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	<b>Control N</b> LOT 1P42C01	78.8 - 131.3 %	99.3 %	OK
	<b>Control A</b> LOT 3P51C00	27.8 - 46.3 %	35.3 %	OK

Prepared by

**Marcin MANUGIEWICZ**

**04. DEZ. 2025**

Date

Approved by

**Alessandro BENIGNI**

**04-Dez-2025**

Date