

CERTIFICATE OF ANALYSIS

Imidazole Buffer 25 mL



Attachment 09 DQ9J0A.02

9J23CC0.02

Temp.: T19J26 07

REF 5410008-RUO
LOT 9J23CC0.02
2027-12-31

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	white		confirmed	OK
	Vial type	white plastic vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Appearance	clear liquid		confirmed	OK
Recovery of controls Fibrinogen	Control N LOT 1P11C00	2.70	-	4.06 g/L	3.87 g/L OK
	Control A LOT 3P23C00	0.88	-	1.32 g/L	1.23 g/L OK

Prepared by

Maylis RAUSCHER

24. MAY 2023

Date

Approved by

Lucas HOLOWICZ

24. MAY 2023

Date