CERTIFICATE OF ANALYSIS
Imidazole Buffer 25 mL

technoclone

Attachment 09 DQ9J0A.02

Temp.: TF0326.07

9J23CC0.02

REF 5410008-RUO

LOT 9J23CC0.02

⌛ 2027-12-31

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	white	confirmed	OK
	Vial type	white plastic vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance	clear liquid	confirmed	OK
Recovery of controls Fibrinogen	**Control N** LOT 1P11C00	2.70 - 4.06 g/L	3.87 g/L	OK
	Control A LOT 3P23C00	0.88 - 1.32 g/L	1.23 g/L	OK

Prepared by Maylis RAUSCHER — 24. MAI 2023 (Date)

Approved by Lucas HOLOVICZ — 24. MAI 2023 (Date)