

# CERTIFICATE OF ANALYSIS

## Imidazole Buffer 25 mL



Attachment 09     DQ9J0A.02

9J23CC0.02

Temp.: TP0326.07

REF	5410008-RUO
LOT	9J23CC0.02
⏳	2027-12-31

Analytical Procedures		Requirement				Result		
visual inspection	Cap color	white				confirmed	OK	
	Vial type	white plastic vial				confirmed	OK	
	Vial label	according to packaging SOP				confirmed	OK	
	Vial barcode	according to packaging SOP				confirmed	OK	
	Box label	according to packaging SOP				confirmed	OK	
	Box barcode	according to packaging SOP				confirmed	OK	
	Package insert	present				confirmed	OK	
	Appearance	clear liquid				confirmed	OK	
Recovery of controls Fibrinogen	<b>Control N</b>							
	LOT 1P11C00	2.70	-	4.06	g/L	3.87	g/L    OK	
<b>Control A</b>								
LOT 3P23C00	0.88	-	1.32	g/L	1.23	g/L    OK		

Prepared by	<div style="text-align: center;"> <p><b>Maylis RAUSCHER</b></p> </div>	<div style="text-align: center;"> <p>24. MAI 2023</p> <p>Date</p> </div>
Approved by	<div style="text-align: center;"> <p><b>Lucas H. LOVICZ</b></p> </div>	<div style="text-align: center;"> <p>24. MAI 2023</p> <p>Date</p> </div>