

CERTIFICATE OF ANALYSIS Imidazole Buffer 25 mL



Attachment 09 DQ9J0A.02

Temp.: TPO326.07

9J23CC0.02

REF 5410008-RUO
LOT 9J23CC0.02
EXP 2027-12-31

Analytical Procedures		Requirement			Result	
visual inspection	Cap color	white			confirmed	OK
	Vial type	white plastic vial			confirmed	OK
	Vial label	according to packaging SOP			confirmed	OK
	Vial barcode	according to packaging SOP			confirmed	OK
	Box label	according to packaging SOP			confirmed	OK
	Box barcode	according to packaging SOP			confirmed	OK
	Package insert	present			confirmed	OK
	Appearance	clear liquid			confirmed	OK
Recovery of controls Fibrinogen	Control N LOT 1P11C00	2.70	-	4.06 g/L	3.87 g/L	OK
	Control A LOT 3P23C00	0.88	-	1.32 g/L	1.23 g/L	OK

Prepared by

Maylis RAUSCHER

24. MAI 2023

Date

Approved by

Lucas HOLEVICZ

24. MAI 2023

Date