

CERTIFICATE OF ANALYSIS
Imidazole Buffer 90 mL



Attachment 12 DQ9J0A.02

9J43CE0.02

Temp.: TP026.07

REF 5410012-RUO
LOT 9J43CE0.02
⌚ 2029-10-31

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	white		confirmed	OK
	Vial type	white plastic vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Appearance	clear liquid		confirmed	OK
Recovery of controls Fibrinogen	Control N LOT 1P31C01	2.67	- 4.01 g/L	3.19 g/L	OK
	Control A LOT 3P42C00	0.93	- 1.39 g/L	1.24 g/L	OK

Prepared by

Smilja DURANOVIC
Smilja Duranovic

11. DEZ. 2024

Date

Approved by

Lucas HOLLOVICH
Lucas Hollovich

12. DEZ. 2024

Date