

CERTIFICATE OF ANALYSIS
Imidazole Buffer 90 mL



Attachment 12 DQ910A.02

Temp.: TFO326.07

9J43CE0.02

REF	5410012-RUO
LOT	9J43CE0.02
⌚	2029-10-31

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	white	confirmed	OK
	Vial type	white plastic vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance	clear liquid	confirmed	OK
Recovery of controls Fibrinogen	Control N LOT 1P31C01	2.67 - 4.01 g/L	3.19 g/L	OK
	Control A LOT 3P42C00	0.93 - 1.39 g/L	1.24 g/L	OK

Prepared by Smilja DURANOVIC
Smilja Duranovic

11. DEZ. 2024
Date

Approved by LUCAS HOLJEVIC
Lucas Holjevic

12. DEZ. 2024
Date