

CERTIFICATE OF ANALYSIS  
Imidazole Buffer 90 mL



Attachment 12      DQ9J0A.02  
9J43CE0.02

Temp.: TF0326.07

REF      5410012-RUO  
LOT      9J43CE0.02  
      2029-10-31

Analytical Procedures		Requirement		Result	
visual inspection	Cap color	white		confirmed	OK
	Vial type	white plastic vial		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Vial barcode	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Appearance	clear liquid		confirmed	OK
Recovery of controls Fibrinogen	<b>Control N</b> LOT 1P31C01	2.67	- 4.01 g/L	3.19 g/L	OK
	<b>Control A</b> LOT 3P42C00	0.93	- 1.39 g/L	1.24 g/L	OK

Prepared by      Smilja DURANOVIC

11. DEZ. 2024  
Date

Approved by      Lucas HOLOVIC

12. DEZ. 2024  
Date