CERTIFICATE OF ANALYSIS
Imidazole Buffer 90 mL

technoclone

Attachment 12 DQ9J0A.02

Temp.: TF0326.07

9J31CE0.04

REF 5410012-RUO

LOT 9J31CE0.04

⌛ 2028-04-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	white	confirmed	OK
	Vial type	white plastic vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance	clear liquid	confirmed	OK
Recovery of controls Fibrinogen	**Control N** LOT 1A32C00	2.40 - 3.70 g/L	3.47 g/L	OK
	Control A LOT 3A32C00	0.92 - 1.38 g/L	1.15 g/L	OK

Prepared by Maylis RAUSCHER 23. APR. 2024
Date

Approved by Michael KAFKA 23. APR. 2024
Date