

**CERTIFICATE OF ANALYSIS**  
**Imidazole Buffer 90 mL**



Attachment 12      DQ9J0A.02

Temp.: 110326.07

9J31CE0.05

|     |             |
|-----|-------------|
| REF | 5410012-RUO |
| LOT | 9J31CE0.05  |
|     | 2028-04-30  |

| Analytical Procedures              |                                 | Requirement                | Result    |    |
|------------------------------------|---------------------------------|----------------------------|-----------|----|
| visual inspection                  | Cap color                       | white                      | confirmed | OK |
|                                    | Vial type                       | white plastic vial         | confirmed | OK |
|                                    | Vial label                      | according to packaging SOP | confirmed | OK |
|                                    | Vial barcode                    | according to packaging SOP | confirmed | OK |
|                                    | Box label                       | according to packaging SOP | confirmed | OK |
|                                    | Box barcode                     | according to packaging SOP | confirmed | OK |
|                                    | Package insert                  | present                    | confirmed | OK |
|                                    | Appearance                      | clear liquid               | confirmed | OK |
| Recovery of controls<br>Fibrinogen | <b>Control N</b><br>LOT 1A32C00 | 2.40 - 3.70 g/L            | 3.47 g/L  | OK |
|                                    | <b>Control A</b><br>LOT 3A32C00 | 0.92 - 1.38 g/L            | 1.15 g/L  | OK |

|             |                           |                              |
|-------------|---------------------------|------------------------------|
| Prepared by | <u>Stefan BRUCKBERGER</u> | <u>02. OKT. 2024</u><br>Date |
| Approved by | <u>Fikret NASUFI</u>      | <u>03. OKT. 2024</u><br>Date |