



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRA00AZ

TECHNOZYM vWF:Ag ELISA Kit 96T

REF 5450201-RUO

LOT RA51C00.02

⌚ 2026-03-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:Ag Plate	RNF41C0	2026-03-31
1 x 0.3mL	VWF:Ag Conjugate	RNH41C0	2026-03-31
1x 80mL	Wash buffer Concentrate	RNA42C0	2026-12-31
1 x 90mL	Incubation buffer	RNB42C0	2026-10-31
1 x 12mL	TMB Substrate	RKN51C0	2027-01-31
1 x 12mL	Stop solution	RKK42C0	2026-09-30
1 x 0.5mL	vWF:Ag Cal 1	1NG41C0	2026-04-30
1 x 0.5mL	vWF:Ag Cal 2	2NG41C0	2026-04-30
1 x 0.5mL	vWF:Ag Cal 3	3NG41C0	2026-04-30
1 x 0.5mL	vWF:Ag Cal 4	4NG41C0	2026-04-30
1 x 0.5mL	vWF:Ag Cal 5	5NG41C0	2026-04-30
1 x 0.5mL	vWF:Ag High Con	ANG41C0	2026-04-30
1 x 0.5mL	vWF:Ag Low Con	BNG41C0	2026-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.45	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (ANG41C0)	0.54 - 0.90 IU/mL	0.65	OK
	Low Control (BNG41C0)	0.20 - 0.46 IU/mL	0.30	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

03. JULI 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOLZ
Lucas HOLZ

03. JULI 2025

Datum / Date: