



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRA00AZ

TECHNOZYM vWF:Ag ELISA Kit 96T

REF 5450201-RUO

LOT RA61C00.02

🕒 2027-05-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x	vWF:Ag Plate	RNF51C0	2027-05-31
1 x 0.3mL	VWF:Ag Conjugate	RNH51C0	2027-06-30
1x 80mL	Wash buffer Concentrate	RNA52C0	2027-11-30
1 x 90mL	Incubation buffer	RNB52C0	2027-11-30
1 x 12mL	TMB Substrate	RKN53C0	2027-12-31
1 x 12mL	Stop solution	RKK61C0	2028-03-31
1 x 0.5mL	vWF:Ag Cal 1	1NG41C2	2028-03-31
1 x 0.5mL	vWF:Ag Cal 2	2NG61C0	2028-03-31
1 x 0.5mL	vWF:Ag Cal 3	3NG61C0	2028-03-31
1 x 0.5mL	vWF:Ag Cal 4	4NG61C0	2028-03-31
1 x 0.5mL	vWF:Ag Cal 5	5NG61C0	2028-03-31
1 x 0.5mL	vWF:Ag High Con	ANG61C0	2028-03-31
1 x 0.5mL	vWF:Ag Low Con	BNG61C0	2028-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.41	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (ANG61C0)	0.52 - 0.86 IU/mL	0.79	OK
	Low Control (BNG61C0)	0.16 - 0.38 IU/mL	0.28	OK

Durchführung von
Prepared by

Fikret NASUFI Fikret Nasufi

14. APR. 2026

Datum / Date:

Freigabe von
Approved by

Lukas GRUBELNIG

14. APR. 2026

Datum / Date: