



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRB0.04

Att.09

RB62C00.01

## TECHNOZYM vWF:CBA Collagen Typ III ELISA Kit 96T.

REF	5450301-RUO
LOT	RB62C00.01
	2027-02-28

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND54C0	2027-04-30
1x	0.5 mL	vWF Calibrator 2	2ND54C0	2027-04-30
1x	0.5 mL	vWF Calibrator 3	3ND53C0	2027-04-30
1x	0.5 mL	vWF Calibrator 4	4ND53C0	2027-04-30
1x	0.5 mL	vWF Calibrator 5	5ND51C0	2027-02-28
1x	0.5 mL	vWF High Control	AND56C0	2027-08-31
1x	0.5 mL	vWF Low Control	BND56C0	2027-08-31
1x	12 mL	TMB Substrate	RKN53C0	2027-12-31
1x	12 mL	Stop solution	RKK52C0	2027-07-31
1x	80 mL	Wash buffer concentrate	RNA52C0	2027-11-30
1x	90 mL	Incubation buffer	RNB52C0	2027-11-30
1x	-	vWF:CBA Plate Collagen Typ III	RNC61C0	2028-03-31
1x	0.3 mL	CBA Conjugate	RNE53C0	2027-11-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Cal1-Cal5	≥ 0.7	1.5	OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND56C0	0.46 - 0.76 IU/mL	0.64 IU/mL	OK
	vWF Low Control LOT BND56C0	0.16 - 0.36 IU/mL	0.28 IU/mL	OK



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRB0.04

Att.09

RB62C00.01

Durchgeföhrt von

Prepared by

  
Lucas HOLOVICZ

18. MRZ. 2026

Datum / Date:

Freigabe von

Approved by

  
Lukas GRUBELNIG

18. MRZ. 2026

Datum / Date: