

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQRB0 04
		Att.07
		RB4AC00.02

## TECHNOZYM vWF:CBA Collagen Typ III ELISA Kit 96T.

REF	5450301-RUO
LOT	RB4AC00.02
	2026-07-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND32C1	2026-10-31
1x	0.5 mL	vWF Calibrator 2	2ND31C1	2026-10-31
1x	0.5 mL	vWF Calibrator 3	3ND31C1	2026-10-31
1x	0.5 mL	vWF Calibrator 4	4ND31C1	2026-10-31
1x	0.5 mL	vWF Calibrator 5	5ND31C1	2026-10-31
1x	0.5 mL	vWF High Control	AND32C1	2026-10-31
1x	0.5 mL	vWF Low Control	BND31C1	2026-10-31
1x	12 mL	TMB Substrate	RKN42C0	2026-08-31
1x	12 mL	Stop solution	RKK42C0	2026-09-30
1x	80 mL	Wash buffer concentrate	RNA41C0	2026-08-31
1x	90 mL	Incubation buffer	RNB41C0	2026-07-31
1x	-	vWF:CBA Plate Collagen Typ III	RNC44C0	2026-08-31
1x	0.3 mL	CBA Conjugate	RNE41C0	2026-07-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Ca1-Ca5	≥ 0.7	1.6	OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND32C1	0.72 - 1.20 IU/mL	0.93 IU/mL	OK
	vWF Low Control LOT BND31C1	0.25 - 0.57 IU/mL	0.39 IU/mL	OK



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRB0.04

Att.07

RB4AC00.02

Durchgeföhrt von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

31. OKT. 2024

Datum / Date:

Freigabe von  
Approved by

*Lucas HOLKOWICZ*

31. OKT. 2024

Datum / Date: