



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRP0.02  
Att.16  
RP4CC00.02

## TECHNOZYM ADAMTS13 INH ELISA 96 T.

REF 5450401-RUO  
 LOT RP4CC00.02  
 ⌚ 2026-07-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	⌚
1x	-	ADAMTS13 INH plate 96 T.	RPS45C0	2026-10-31
1x	0.3 mL	ADAMTS13 INH POX	RPV42C0	2026-10-31
1x	0.5 mL	ADAMTS13 INH Calibrator 1	1PT41C0	2026-09-30
1x	0.5 mL	ADAMTS13 INH Calibrator 2	2PT41C0	2026-09-30
1x	0.5 mL	ADAMTS13 INH Calibrator 3	3PT41C0	2026-09-30
1x	0.5 mL	ADAMTS13 INH Calibrator 4	4PT41C0	2026-09-30
1x	0.5 mL	ADAMTS13 INH Calibrator 5	5PT41C0	2026-09-30
1x	0.5 mL	ADAMTS13 INH High Control	RPY41C0	2026-09-30
1x	0.5 mL	ADAMTS13 INH Low Control	RPZ41C0	2026-09-30
1x	12 mL	Stop Solution	RKK42C0	2026-09-30
1x	12 mL	TMB Substrate	RKN42C0	2026-08-31
1x	80 mL	Wash buffer concentrate	RNA41C0	2026-08-31
1x	90 mL	Incubation buffer	RNB41C0	2026-07-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	<b>Kurvensteigung Slope</b>	≥ 0.7	1.3	OK
	<b>ADAMTS13 INH High Control</b> LOT RPY41C0	41.54 - 69.24 U/mL	58.68 U/mL	OK
	<b>ADAMTS13 INH Low Control</b> LOT RPZ41C0	10.31 - 24.07 U/mL	17.10 U/mL	OK

Durchgeführt von  
Prepared by

*Lidia Holowicz*  
\_\_\_\_\_  
Lidia Holowicz

16. JAN. 2025  
Datum / Date:

Freigabe von  
Approved by

Tikret NASUFI *Tikret Nasufi*  
\_\_\_\_\_  
Tikret Nasufi

16. JAN. 2025  
Datum / Date: