

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP0.02
		Att.18
		6PT51C0.02

TECHNOZYM ADAMTS13 INH Calibrator Set 5x0.5 mL

REF	5450461-RUO
LOT	6PT51C0.02
	2027-04-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	ADAMTS13 INH Calibrator 1	1PT51C0	2027-04-30
1x	0.5 mL	ADAMTS13 INH Calibrator 2	2PT51C0	2027-04-30
1x	0.5 mL	ADAMTS13 INH Calibrator 3	3PT51C0	2027-05-31
1x	0.5 mL	ADAMTS13 INH Calibrator 4	4PT51C0	2027-05-31
1x	0.5 mL	ADAMTS13 INH Calibrator 5	5PT51C0	2027-05-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendung Usability	Kurvensteigung Slope	≥ 0.7	1.2	OK
	ADAMTS13 INH High Control LOT RPY51C0	41.89 - 69.81 U/mL	62.59 U/mL	OK
	ADAMTS13 INH Low Control LOT RPZ51C0	10.81 - 25.23 U/mL	17.51 U/mL	OK

Durchgeföhrt von
Prepared by

Fikret NASUFI *Fikret Nasufi*

04. SEP. 2025

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ *Lucas HoloVICZ*

05. SEP. 2025

Datum / Date: