


	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP0.02
		Att.19
		7PT42C0.02

TECHNOZYM ADAMTS13 INH Control Set 2x0.5 mL

REF	5450463-RUO
LOT	7PT42C0.02
	2026-09-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	ADAMTS13 INH High Control	RPY41C0	2026-09-30
1x	0.5 mL	ADAMTS13 INH Low Control	RPZ41C0	2026-09-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Kontrollwiederfindung Recovery of controls	ADAMTS13 INH High Control LOT RPY41C0	41.54 - 69.24 U/mL	62.55 U/mL	OK
	ADAMTS13 INH Low Control LOT RPZ41C0	10.31 - 24.07 U/mL	16.35 U/mL	OK

Durchgeföhrt von Prepared by	 Lucas HOLZNER	09. OKT. 2024 Datum / Date:
Freigabe von Approved by	 Maylis FAUSCHER	09. OKT. 2024 Datum / Date: