

technoclone TC	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP0.02 Att.19 7PT51C0.02
--	--	---

TECHNOZYM ADAMTS13 INH Control Set 2x0.5 mL

REF

LOT

5450463-RUO

7PT51C0.02

2027-05-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	ADAMTS13 INH High Control	RPY51C0	2027-05-31
1x	0.5 mL	ADAMTS13 INH Low Control	RPZ51C0	2027-05-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement		Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
Kontrollwiederfindung Recovery of controls	ADAMTS13 INH High Control LOT RPY51C0	41.89	- 69.81 U/mL	62.59 U/mL	OK
	ADAMTS13 INH Low Control LOT RPZ51C0	10.81	- 25.23 U/mL	17.51 U/mL	OK

Durchgeföhrt von
Prepared by

Fikret NASUFI

0 4. SEP. 2025

Datum / Date:

Freigabe von
Approved by

Lucas GOLOVICZ

0 5. SEP. 2025

Datum / Date: