


CERTIFICATE OF ANALYSIS
TECHNOZYM ADAMTS13 INH CONT Set



Temp.: TP0327.07

Attachment 19 DQRP00.03

7PT61C0.02

REF	5450463-RUO
LOT	7PT61C0.02
	2028-02-29

Content	volume	Reagent	LOT	
1x	0.5 mL	ADAMTS13 INH High Control	RPY61C0	2028-02-29
1x	0.5 mL	ADAMTS13 INH Low Control	RPZ61C0	2028-02-29

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	ADAMTS13 INH High Control LOT RPY61C0	35.97 - 59.95 U/mL	54.30 U/mL	OK
	ADAMTS13 INH Low Control LOT RPZ61C0	8.21 - 19.15 U/mL	13.92 U/mL	OK

Prepared by Lucas GLOVICZ

23. MRZ. 2026
Date

Approved by Fikret NASUFI

24. MRZ. 2026
Date