



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRP0.02
Att.19
7PT41C0.02

TECHNOZYM ADAMTS13 INH Control Set 2x0.5 mL

| | |
|-----|-------------|
| REF | 5450463-RUO |
| LOT | 7PT41C0.02 |
| | 2025-11-30 |

| Inhalt Content | Volumen volume | Reagenz Reagent | LOT | |
|-------------------|-------------------|---------------------------|---------|------------|
| 1x | 0.5 mL | ADAMTS13 INH High Control | RPY31C0 | 2025-11-30 |
| 1x | 0.5 mL | ADAMTS13 INH Low Control | RPZ31C0 | 2025-11-30 |

| Analytische Verfahren Analytical Procedures | Anforderung Requirement | Ergebnis Result |
|--|--|----------------------------|
| visuelle Kontrolle visual inspection | Content | according to packaging SOP |
| | Vial label | confirmed |
| | Box label | confirmed |
| | Box barcode | confirmed |
| | Package insert | confirmed |
| | Batch table | confirmed |
| Kontrollwiederfindung Recovery of controls | ADAMTS13 INH High Control LOT RPY31C0 | 42.53 - 70.88 U/mL |
| | ADAMTS13 INH Low Control LOT RPZ31C0 | 59.43 U/mL |
| | 13.17 - 30.73 U/mL | 22.10 U/mL |

Durchgeführt von
Prepared by

Fikret NASUFI *Fikret Nasu*

16. FEB. 2024

Datum / Date:

Freigabe von
Approved by

Lucas HOLZ *Lucas Holz*

16. FEB. 2024

Datum / Date: