

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP0.02 Att.19 7PT41C0.02
---	---	--

TECHNOZYM ADAMTS13 INH Control Set 2x0.5 mL

REF	5450463-RUO
LOT	7PT41C0.02
	2025-11-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	ADAMTS13 INH High Control	RPY31C0	2025-11-30
1x	0.5 mL	ADAMTS13 INH Low Control	RPZ31C0	2025-11-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Content	according to packaging SOP confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Box barcode	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Batch table	present confirmed OK
Kontrollwiederfindung Recovery of controls	ADAMTS13 INH High Control LOT RPY31C0	42.53 - 70.88 U/mL 59.43 U/mL OK
	ADAMTS13 INH Low Control LOT RPZ31C0	13.17 - 30.73 U/mL 22.10 U/mL OK

Durchgeführt von
Prepared by

Fikret NASUFI *Fikret Nasu*

16. FEB. 2024

Datum / Date:

Freigabe von
Approved by

Lucas HOLZ *Lucas Holz*

16. FEB. 2024

Datum / Date: