

CERTIFICATE OF ANALYSIS

TECHNOZYM ADAMTS13 Antigen CAL Set



Attachment 13 DQxRT0.05

6RT52C0.02

Temp.: TFC027.07

REF 5450661-RUO
 LOT 6RT52C0.02
 2027-01-31

Content	Volume	Reagent	LOT	
1x	0.5 mL	ADAMTS13 Antigen CAL 1	1RT42C1	2027-07-31
1x	0.5 mL	ADAMTS13 Antigen CAL 2	2RT51C0	2027-01-31
1x	0.5 mL	ADAMTS13 Antigen CAL 3	3RT51C0	2027-01-31
1x	0.5 mL	ADAMTS13 Antigen CAL 4	4RT51C0	2027-01-31
1x	0.5 mL	ADAMTS13 Antigen CAL 5	5RT51C0	2027-01-31

Analytical Procedures		Requirement		Result	
visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
Usability	delta OD Cal 1 - Cal 5	≥ 0.7		0.7	OK
Recovery of controls	ADAMTS13 Antigen CONT H LOT RRY51C0	0.59	-	0.99 IU/mL	0.91 IU/mL OK
	ADAMTS13 Antigen CONT L LOT RRZ51C0	0.09	-	0.21 IU/mL	0.18 IU/mL OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

06. NOV. 2025

Date

Approved by

Lukas Grubelnig
 Lukas GRUBELNIG

06. NOV. 2025

Date