

CERTIFICATE OF ANALYSIS

TECHNOZYM ADAMTS13 Antigen CONT Set



Attachment 14 DQxRT0.05

7RT51C0.03

Temp.: TP0327.07

REF 5450663-RUO
 LOT 7RT51C0.03
 ⌚ 2027-01-31

Content	volume	Reagent	LOT	⌚
1x	0.5 mL	ADAMTS13 Antigen CONT H	RRY51C0	2027-01-31
1x	0.5 mL	ADAMTS13 Antigen CONT L	RRZ51C0	2027-01-31

Analytical Procedures		Requirement		Result	
visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
Recovery of controls	ADAMTS13 Antigen CONT H	0.59	- 0.99 IU/mL	0.89 IU/mL	OK
	ADAMTS13 Antigen CONT L	0.09	- 0.21 IU/mL	0.15 IU/mL	OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

06. NOV. 2025

Date

Approved by

LU
 Lukas GRUBELNIG

06. NOV. 2025

Date