


**CERTIFICATE OF ANALYSIS**  
**TECHNOZYM ADAMTS13 Activity ELISA Kit**




Attachment 03      DQRU00.11

Temp.: TPO327.07

RU63C00.03

REF            5450701-RUO  
 LOT            RU63C00.03  
            2027-08-31

Content	volume	Reagent	LOT	
2x	6mL	ADAMTS-13 Activity vWF73 Substrate	605	2027-08-31
1x	30mL	ADAMTS-13 Activity Reaction Buffer	605	2027-08-31
1x	12mL	ADAMTS-13 Activity HRP Conj Antibody	605	2027-08-31
1x	12mL	ADAMTS-13 Activity Color Reagent	605	2027-08-31
1x	53mL	ADAMTS-13 Activity Wash Buffer Conc	605	2027-08-31
1x	12mL	ADAMTS-13 Activity Stop Solution	605	2027-08-31
1x	n.a.	ADAMTS-13 Activity Plate 96T.	605	2027-08-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 1	1UT52C0	2027-09-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 2	2UT52C0	2027-09-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 3	3UT52C0	2027-09-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 4	4UT52C0	2027-09-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 6	6UT52C0	2027-09-30
1x	0.5mL	UY ADAMTS-13 Activity High Control	RUY52C0	2027-09-30
1x	0.5mL	UZ ADAMTS-13 Activity Low Control	RUZ52C0	2027-09-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 5	5UT52C0	2027-09-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	<b>deltaOD Cal1-Cal6</b>	≥ 0.7	1.5	OK
Recovery of controls	<b>ADAMTS-13 Activity High Control</b> LOT RUY52C0	0.68 - 1.09 IU/mL	0.80 IU/mL	OK
	<b>ADAMTS-13 Activity Low Control</b> LOT RUZ52C0	0.13 - 0.31 IU/mL	0.23 IU/mL	OK

Prepared by Lucas HOLOVICZ

22. FEB. 2026  
Date

Approved by Lukas GRUBELNIG

23. APR. 2026  
Date