


**CERTIFICATE OF ANALYSIS**  
**TECHNOZYM ADAMTS13 Activity CAL Set**




Temp.: TP0327.07

Attachment 06      DQxUT0.04

8UT31C0.02

REF      5450761-RUO  
 LOT      8UT31C0.02  
      2025-05-31

Content	Volume	Reagent	LOT	
1x	0.5 mL	ADAMTS13 Activity Calibrator 1	1UT31C0	2025-05-31
1x	0.5 mL	ADAMTS13 Activity Calibrator 2	2UT31C0	2025-05-31
1x	0.5 mL	ADAMTS13 Activity Calibrator 3	3UT31C0	2025-05-31
1x	0.5 mL	ADAMTS13 Activity Calibrator 4	4UT31C0	2025-05-31
1x	0.5 mL	ADAMTS13 Activity Calibrator 5	5UT31C0	2025-05-31
1x	0.5 mL	ADAMTS13 Activity Calibrator 6	6UT31C0	2025-05-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Usability	delta OD Cal 1 - Cal 6	≥ 0.7	1.6	OK
Recovery of controls	<b>ADAMTS13 Activity CONT H</b> LOT RUY31C0	0.71 - 1.19 IU/mL	1.02 IU/mL	OK
	<b>ADAMTS13 Activity CONT L</b> LOT RUZ31C0	0.17 - 0.39 IU/mL	0.32 IU/mL	OK

Prepared by

Michael KAFKA 

03. AUG. 2023

Date

Approved by

  
 Suzana GRESAK PODAKAR

03. AUG. 2023

Date