

CERTIFICATE OF ANALYSIS
TECHNOZYM ADAMTS13 Activity CONT Set



Attachment 08 DQxUT0.05

7UT42C0.02

Temp.: TFO327.07

REF 5450763-RUO
LOT 7UT42C0.02
⌚ 2026-08-31

Content	volume	Reagent	LOT	⌚
1x	0.5 mL	ADAMTS13 Activity High Control	RUY42C0	2026-08-31
1x	0.5 mL	ADAMTS13 Activity Low Control	RUZ42C0	2026-08-31

Analytical Procedures		Requirement		Result	
visual inspection	Content	according to packaging SOP		confirmed	OK
	Vial label	according to packaging SOP		confirmed	OK
	Box label	according to packaging SOP		confirmed	OK
	Box barcode	according to packaging SOP		confirmed	OK
	Package insert	present		confirmed	OK
	Batch table	present		confirmed	OK
Recovery of controls	ADAMTS13 Activity CONT H	0.71	- 1.19 IU/mL	1.00 IU/mL	OK
	ADAMTS13 Activity CONT L	0.17	- 0.39 IU/mL	0.29 IU/mL	OK

Prepared by

Fikret NASUFI *Fikret Nasufi*

22. OKT. 2024

Date

Approved by

Lukas HOLYICZ
Lukas HOLYICZ

22. OKT. 2024

Date