

**CERTIFICATE OF ANALYSIS**  
**TECHNOZYM ADAMTS13 Activity CONT Set**



Attachment 08      DQxUT0.05

Temp.: TF0127.07

7UT42C0.02

REF	5450763-RUO
LOT	7UT42C0.02
⌚	2026-08-31

Content	volume	Reagent	LOT	⌚
1x	0.5 mL	ADAMTS13 Activity High Control	RUY42C0	2026-08-31
1x	0.5 mL	ADAMTS13 Activity Low Control	RUZ42C0	2026-08-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Recovery of controls	<b>ADAMTS13 Activity CONT H</b>	0.71 - 1.19 IU/mL	1.00 IU/mL	OK
	<b>ADAMTS13 Activity CONT L</b>	0.17 - 0.39 IU/mL	0.29 IU/mL	OK

Prepared by Fikret NASUFI *Fikret Nasufi*

22. OKT. 2024

Date

Approved by *[Signature]*

22. OKT. 2024

Date