

CERTIFICATE OF ANALYSIS

Technofluor ADAMTS13:Act



Attachment 06 DQFA00.02

FA51C00.02

Temp.: TF0327 07

REF 5800100-RUO
 LOT FA51C00.02
 2027-01-31

Content	volume	Reagent	LOT	
2x	3mL	Technofluor ADAMTS13 BUF	FAB51C0	2027-01-31
2x	1mL	Technofluor ADAMTS13 SUB	FAA51C0	2027-01-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N			
	LOT 1P31C01	0.65 - 1.08 IU/mL	0.86 IU/mL	OK
	Control A			
	LOT 3P42C00	0.32 - 0.54 IU/mL	0.42 IU/mL	OK

Prepared by

Smilja DURANOVIC
Duranovic

21. FEB. 2025

Date

Approved by

Heidi HOFMEIER

21. FEB. 2025

Date