

CERTIFICATE OF ANALYSIS

Technofluor ADAMTS13:Act



Attachment 06 DQFA00.02

FA51C00.02

Temp.: TF0327.07

REF 5800100-RUO
 LOT FA51C00.02
 2027-01-31

| Content | volume | Reagent | LOT | |
|---------|--------|--------------------------|---------|------------|
| 2x | 3mL | Technofluor ADAMTS13 BUF | FAB51C0 | 2027-01-31 |
| 2x | 1mL | Technofluor ADAMTS13 SUB | FAA51C0 | 2027-01-31 |

| Analytical Procedures | | Requirement | Result | |
|-----------------------|------------------|----------------------------|------------|----|
| visual inspection | Content | according to packaging SOP | confirmed | OK |
| | Vial label | according to packaging SOP | confirmed | OK |
| | Vial barcode | according to packaging SOP | confirmed | OK |
| | Box label | according to packaging SOP | confirmed | OK |
| | Box barcode | according to packaging SOP | confirmed | OK |
| | Package insert | present | confirmed | OK |
| Recovery of controls | Control N | | | |
| | LOT 1P31C01 | 0.65 - 1.08 IU/mL | 0.86 IU/mL | OK |
| | Control A | | | |
| | LOT 3P42C00 | 0.32 - 0.54 IU/mL | 0.42 IU/mL | OK |

Prepared by

Smilja DURANOVIĆ
 Duranovic

21. FEB. 2025

Date

Approved by

Mark HOLBOURN
 Holboorn

21. FEB. 2025

Date