

CERTIFICATE OF ANALYSIS
Technofluor ADAMTS13:Act



Temp.: TP0327.07

Attachment 06 DQFA00.02

FA54C00.02


REF	5800100-RUO
LOT	FA54C00.02
⌚	2027-11-30

Content	volume	Reagent	LOT	⌚
2x	3mL	Technofluor ADAMTS13 BUF	FAB54C0	2027-11-30
2x	1mL	Technofluor ADAMTS13 SUB	FAA54C0	2027-11-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N LOT 1P53C00	0.76 - 1.26 IU/mL	0.97 IU/mL	OK
	Control A LOT 3P54C00	0.35 - 0.58 IU/mL	0.47 IU/mL	OK

Prepared by 
Lukas GRUBELNIG

18. DEZ. 2025
 Date

Approved by 
Alessandro BENIGNI

18-Dez-2025
 Date