

# CERTIFICATE OF ANALYSIS

## Technofluor ADAMTS13:Act



Attachment 06 DQFA00.02

Temp.: TF0327.07

FA43C00.02

REF	5800100-RUO
LOT	FA43C00.02
	2026-04-30

Content	volume	Reagent	LOT	
2x	3mL	Technofluor ADAMTS13 BUF	FAB42C0	2026-04-30
2x	1mL	Technofluor ADAMTS13 SUB	FAA42C0	2026-04-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	<b>Control N</b> LOT 1P31C00	0.65 - 1.08 IU/mL	0.80 IU/mL	OK
	<b>Control A</b> LOT 3P23C01	0.31 - 0.51 IU/mL	0.42 IU/mL	OK

Prepared by Lucas HOLOWICZ

29. MAI 2024  
Date

Approved by Michael KAFKA

29. MAI 2024  
Date