

**CERTIFICATE OF ANALYSIS**  
**Technofluor ADAMTS13:Act**



Temp.: TP0327.07

Attachment 06      DQFA00.02

FA43C00.04

REF	5800100-RUO
LOT	FA43C00.04
	2026-04-30

Content	volume	Reagent	LOT	
2x	3mL	Technofluor ADAMTS13 BUF	FAB42C0	2026-04-30
2x	1mL	Technofluor ADAMTS13 SUB	FAA42C0	2026-04-30

Analytical Procedures	Requirement	Result
visual inspection	Content	according to packaging SOP      confirmed      OK
	Vial label	according to packaging SOP      confirmed      OK
	Vial barcode	according to packaging SOP      confirmed      OK
	Box label	according to packaging SOP      confirmed      OK
	Box barcode	according to packaging SOP      confirmed      OK
	Package insert	present      confirmed      OK
Recovery of controls	<b>Control N</b> LOT 1P31C01	0.65      -      1.08 IU/mL      0.89 IU/mL      OK
	<b>Control A</b> LOT 3P42C00	0.32      -      0.54 IU/mL      0.40 IU/mL      OK

Prepared by Smilja DURANOVIĆ  
*Duranovic*

13. SEP. 2024  
Date

Approved by Mavis FAUSCHER  
*M. Faucher*

13. SEP. 2024  
Date