

CERTIFICATE OF ANALYSIS

Technofluor ADAMTS13:Act



Temp.: TF0327.07

Attachment 06 DQFA00.02

FA44C00.02

REF	5800100-RUO
LOT	FA44C00.02
	2026-08-31

Content	volume	Reagent	LOT	
2x	3mL	Technofluor ADAMTS13 BUF	FAB43C0	2026-08-31
2x	1mL	Technofluor ADAMTS13 SUB	FAA43C0	2026-09-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N LOT 1P31C01	0.65 - 1.08 IU/mL	0.81 IU/mL	OK
	Control A LOT 3P42C00	0.32 - 0.54 IU/mL	0.42 IU/mL	OK

Prepared by Smilja DURANOVIC
Smilja Duranovic

02. OKT. 2024

Date

Approved by Lucas HOLWIGZ
Lucas Holwigz

03. OKT. 2024

Date