

# CERTIFICATE OF ANALYSIS


## Technofluor ADAMTS13:Act




Attachment 06      DQFA00.02

Temp.: TF0327.07

FA52C00.02

REF	5800100-RUO
LOT	FA52C00.02
	2027-05-31

Content	volume	Reagent	LOT	
2x	3mL	Technofluor ADAMTS13 BUF	FAB52C0	2027-05-31
2x	1mL	Technofluor ADAMTS13 SUB	FAA52C0	2027-05-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	<b>Control N</b> LOT 1P42C00	0.76 - 1.26 IU/mL	0.94 IU/mL	OK
	<b>Control A</b> LOT 3P42C01	0.32 - 0.54 IU/mL	0.43 IU/mL	OK

Prepared by

  
Lavdrim NASUFI

10. JULI 2025

Date

Approved by

  
Lucas HOLOVICZ

11. JULI 2025

Date