

CERTIFICATE OF ANALYSIS

Technofluor ADAMTS13:Act 0 Cal



Attachment 07 DQFA00.02

Temp.: TFS326.07

FAC42C0.03

REF 5800102-RUO
LOT FAC42C0.03
2026-09-30

Analytical Procedures			Requirement		Result		
visual inspection	Cap color		grey		confirmed	OK	
	Vial type		clear glass vial		confirmed	OK	
	Vial label		according to packaging SOP		confirmed	OK	
	Vial barcode		according to packaging SOP		confirmed	OK	
	Box label		according to packaging SOP		confirmed	OK	
	Box barcode		according to packaging SOP		confirmed	OK	
	Package insert		present		confirmed	OK	
	Appearance before reconstitution		buff colored plug of lyophilized material		confirmed	OK	
Appearance after reconstitution		amber colored liquid		confirmed	OK		
Virology			HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab		Negative Negative Negative		OK
The testing methods applied were FDA-approved or CE marked.							
Recovery of controls	Control N						
	LOT	1P31C01	0.65	-	1.08 IU/mL	0.89 IU/mL	OK
	Control A						
	LOT	3P42C00	0.32	-	0.54 IU/mL	0.43 IU/mL	OK

Prepared by

Stefan BRUCKBERGER

27. MRZ. 2025

Date

Approved by

Smilja DURANOVIC

28. MRZ. 2025

Date