

CERTIFICATE OF ANALYSIS
Technofluor ADAMTS13:Act 0 Cal



Temp.: TF0316.07

Attachment 07 DQFA00.02

FAC53C0.02

REF 5800102-RUO
 LOT FAC53C0.02
 ⌚ 2027-11-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	green	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	amber colored liquid	confirmed	OK
Virology	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
The testing methods applied were FDA-approved or CE marked.				
Recovery of controls	Control N LOT 1P53C00	0.76 - 1.26 IU/mL	0.93 IU/mL	OK
	Control A LOT 3P54C00	0.35 - 0.58 IU/mL	0.47 IU/mL	OK

Prepared by Alessandro BENIGNI

22-Jan-2026
Date

Approved by Lukas GRUBELNIG

22-Jan-2026
Date