

CERTIFICATE OF ANALYSIS TECHNOFLUOR FXIII Activity



Attachment 06 DQFB00.02

Temp.: T0327.07

FB42C00.02

REF	5800200-RUO
LOT	FB42C00.02
⌚	2026-09-30

Content	volume	Reagent	LOT	⌚
2x	3mL	HFP: Technofluor FXIII Buffer 3mL	FBC42C0	2026-09-30
2x	2mL	HFP: Technofluor FXIII Substrate 2mL	FBB42C0	2026-09-30
2x	2mL	HFP: Technofluor FXIII Reagent 2mL	FBA42C0	2026-09-30

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N LOT 1P31C01	60.0 - 100.0 %	88.9 %	OK
	Control A LOT 3P42C00	28.5 - 47.5 %	35.5 %	OK

Prepared by Smilja DURANOVIC
Smilja Duranovic

27. SEP. 2024

Date

Approved by Martin FAUSCHER
Martin Fauscher

27. SEP. 2024

Date